



LABORATORY ANALYTICAL REPORT

Client: Texas Water Development Board
Project: TWDB
Lab ID: C07070695-001
Client Sample ID: 3240408

Report Date: 08/02/07
Collection Date: 07/09/07 11:00
Date Received: 07/17/07
Matrix: Aqueous

Analyses	Result	Units	Qual	MCL/		Method	Analysis Date / By
				RL	QCL		
MAJOR IONS							
Alkalinity, Phenolphthalein as CaCO3	14	mg/L		1		A2320 B	07/23/07 08:37 / bas
Alkalinity, Total as CaCO3	578	mg/L		1		A2320 B	07/23/07 08:37 / bas
Bromide	ND	mg/L		0.50		E300.0	07/26/07 15:48 / bt
Calcium	2.4	mg/L		0.5		E200.7	07/26/07 13:22 / cp
Chloride	38	mg/L		1		A4500-Cl B	07/18/07 13:59 / jl
Fluoride	6.1	mg/L		0.1		A4500-F C	07/23/07 12:36 / bas
Magnesium	1.1	mg/L		0.5		E200.7	07/26/07 13:22 / cp
Nitrogen, Nitrate+Nitrite as N	0.4	mg/L		0.1		E353.2	07/19/07 12:04 / jal
Potassium	1.8	mg/L		0.5		E200.7	07/26/07 13:22 / cp
Silica	12.8	mg/L		0.1		E200.7	07/26/07 13:22 / cp
Sodium	381	mg/L	D	0.8		E200.7	07/26/07 13:22 / cp
Sulfate	256	mg/L	D	6		A4500-SO4 E	07/17/07 13:44 / zd
METALS - DISSOLVED							
Aluminum	4	ug/L		1		E200.8	07/30/07 03:11 / bws
Antimony	ND	ug/L		1		E200.8	07/30/07 03:11 / bws
Arsenic	ND	ug/L		1		E200.8	07/30/07 03:11 / bws
Barium	7	ug/L		1		E200.8	07/30/07 03:11 / bws
Beryllium	ND	ug/L		1		E200.8	07/30/07 03:11 / bws
Boron	1540	ug/L		100		E200.7	07/26/07 13:22 / cp
Cadmium	ND	ug/L		1		E200.8	07/30/07 03:11 / bws
Chromium	ND	ug/L		1		E200.8	07/30/07 03:11 / bws
Cobalt	ND	ug/L		1		E200.8	07/30/07 03:11 / bws
Copper	3	ug/L		1		E200.8	07/30/07 03:11 / bws
Iron	ND	ug/L		30		E200.7	07/26/07 13:22 / cp
Lead	ND	ug/L		1		E200.8	07/30/07 03:11 / bws
Lithium	65	ug/L		1		E200.8	07/24/07 11:16 / sml
Manganese	2	ug/L		1		E200.8	07/30/07 03:11 / bws
Molybdenum	2	ug/L		1		E200.8	07/30/07 03:11 / bws
Selenium	ND	ug/L		1		E200.8	07/30/07 03:11 / bws
Strontium	239	ug/L		1		E200.8	07/30/07 03:11 / bws
Thallium	ND	ug/L		1		E200.8	07/30/07 03:11 / bws
Vanadium	ND	ug/L		1		E200.8	07/30/07 03:11 / bws
Zinc	4	ug/L		1		E200.8	07/30/07 03:11 / bws
DATA QUALITY							
A/C Balance (± 5)	4.08	%		250		Calculation	07/30/07 14:12 / bws
Anions	18.3	meq/L		250		Calculation	07/30/07 14:12 / bws
Cations	16.9	meq/L		250		Calculation	07/30/07 14:12 / bws

Report Definitions:
 RL - Analyte reporting limit.
 QCL - Quality control limit.
 D - RL increased due to sample matrix interference.

MCL - Maximum contaminant level.
 ND - Not detected at the reporting limit.