

6390 Joyce Drive  
# 100  
Golden, CO 80403

Phone 303-940-0033  
Fax 866-283-0269  
[www.phenova.com](http://www.phenova.com)

August 22, 2019

Ms. Julie Weisz  
Energy Laboratories, Inc.  
400 West Boxelder Road  
Gillette, WY 82718  
USA

Dear Julie,

Thank you for participating in the R25343 Rapid Return™ Proficiency Testing Study. Enclosed is your final report, which has been carefully reviewed by the PT specialists at Phenova. A final report for your laboratory has been sent to all accrediting agencies you requested at the time of your data submittal.

Thank you again for participating in the R25343 Rapid Return™ Proficiency Testing Study. We appreciate working with you and look forward to our next study. If you have any questions, please call us at 866-942-2978.

**Report Definitions:**

***Assigned Value***

The Assigned Value is determined from the study mean, gravimetric and volumetric true concentration of an analyte to be analyzed, calculation and/or an appropriate reference value as stipulated in the EPA National Standards for Water Proficiency Testing Studies Criteria Document (current version), the National Environmental Laboratory Accreditation Conference Institute (TNI) criteria (ref: TNI FOT tables, TNI PT Committee) and other documents distributed by accrediting agencies as applicable.

***Evaluation Limits***

Acceptance Limits are derived from fixed limits, coefficients, constants and calculations stipulated in the EPA National Standards for Water Proficiency Testing Studies Criteria Documents (current version), the National Environmental Laboratory Accreditation Conference Institute (TNI) criteria (ref: TNI FOT tables, TNI PT Committee) and other documents distributed by accrediting agencies as applicable.

***Evaluation***

<i>Acceptable</i>	The reported value falls within the Acceptance Limits.
<i>Not Acceptable</i>	The reported value falls outside the Acceptance Limits.
<i>No Evaluation</i>	The reported value is non-numeric and can not be evaluated.
<i>NR</i>	As required by the TNI standards and requested by state authorities, any analyte purchased but not reported by your facility is listed as NR (Not Reported).

This report must not be reproduced except in full, without the written approval of Phenova.

Various aspects of Phenova's proficiency testing schemes can from time to time be subcontracted. If subcontracting occurs, it is placed with a competent and qualified subcontractor and Phenova is responsible for all work performed.

This report must not be used to claim product certification, approval, or endorsement by the American Association for Laboratory Accreditation (A2LA) or any agency of the federal government.



6390 Joyce Drive  
# 100  
Golden, CO 80403

Phone 303-940-0033  
Fax 866-283-0269  
www.phenova.com

State Regulatory Agency Request(s):

No requests were made to send a copy of your R25343 final results to a state agency.

Other Regulatory Agency Request(s):

No requests were made to send a copy of your R25343 final results to an agency.

Third Party Request(s):

No requests were made to send a copy of your R25343 final results to a third party.

6390 Joyce Drive      Phone 303-940-0033  
# 100                      Fax 866-283-0269  
Golden, CO 80403      www.phenova.com

## Final Report - Rapid Return<sup>TM</sup>

### Study: R25343

**Opening Date: July 19, 2019 - Closing Date: August 20, 2019**

Laboratory: Energy Laboratories, Inc.  
400 West Boxelder Road  
Gillette, WY 82718  
USA

Contact: Ms. Julie Weisz, Quality Assurance Manager  
307-686-7175

EPA Lab ID: WY00006

## Total Organic Halides (RR-TOX-WP) Lot #: R25343-24

NELAC Code	Analyte	Analysis Date	Analyst	Method Code	Method Description	Units	Assigned Value	Result	Acceptance Limits	Evaluation
<b>Additional State Specific Analytes</b>										
2045	Total Organic Halides	8/13/2019	David Bowles	10194408	EPA 9020B 1994	µg/L	803	772	616 - 944	Acceptable